



# eBook for Undergraduate Education in Radiology

| **CHAPTER:** Magnetic Resonance Imaging



### Preface

Undergraduate teaching of radiology in Europe is provided according to national schemes and may vary considerably from one academic institution to another. Sometimes, the field of radiology is considered as a “cross-cutting discipline” or taught within the context of other clinical disciplines, e.g., internal medicine or surgery.

This e-book has been created in order to serve medical students and academic teachers throughout Europe to understand and teach radiology as a whole coherent discipline, respectively. Its contents are based on the *Undergraduate Level of the ESR European Training Curriculum for Radiology* and summarize the so-called **core elements** that may be considered as the basics that every medical student should be familiar with. Although specific radiologic diagnostic skills for image interpretation cannot be acquired by all students and rather belong to the learning objectives of the *Postgraduate Levels of the ESR Training Curricula*, the present e-book also contains some **further insights** related to modern imaging in the form of examples of key pathologies, as seen by the different imaging modalities. These are intended to give the interested undergraduate student an understanding of modern radiology, reflecting its multidisciplinary character as an organ-based specialty.

We would like to extend our special thanks to the authors and members of the ESR Education Committee who have contributed to this eBook, to Carlo Catalano, Andrea Laghi and András Palkó who initiated this project, and to the ESR Office, in particular Bettina Leimberger and Danijel Lepir, for all their support in realising this project.

We hope that this e-book may fulfil its purpose as a useful tool for undergraduate academic radiology teaching.

Minerva Becker

### Chapter Outline

**MRI System**

**MRI Principle**

**The Spin Echo Sequence**

**The Gradient  
Echo Sequence**

**Other Sequences: The  
MRI Jungle**

**Contrast Agents in MRI**

**Artefacts in MRI**

**MRI Advantages  
and Disadvantages**

**Take Home Messages**

**References**

**Test Your Knowledge**





## Copyright and Terms of Use

This work is licensed under a [Creative Commons Attribution-NonCommercial-NoDerivatives 4.0 International License](#).

You are free to:

- **Share** – copy and redistribute the material in any medium or format

Under the following terms:

- **Attribution** – You must give appropriate credit, provide a link to the license, and indicate if changes were made. You may do so in any reasonable manner, but not in any way that suggests the licensor endorses you or your use.
- **NonCommercial** – You may not use the material for commercial purposes.
- **NoDerivatives** – If you remix, transform, or build upon the material, you may not distribute the modified material.

How to cite this work:

European Society of Radiology, Bénédicte MA Delattre, Andrea Laghi, András Kincses, Minerva Becker (2024) eBook for Undergraduate Education in Radiology: Magnetic Resonance Imaging. DOI 10.26044/esr-undergraduate-ebook-24

## Chapter Outline

**MRI System**

**MRI Principle**

**The Spin Echo Sequence**

**The Gradient Echo Sequence**

**Other Sequences: The MRI Jungle**

**Contrast Agents in MRI**

**Artefacts in MRI**

**MRI Advantages and Disadvantages**

**Take Home Messages**

**References**

**Test Your Knowledge**



## Chapter Outline

**MRI System**

**MRI Principle**

**The Spin Echo Sequence**

**The Gradient  
Echo Sequence**

**Other Sequences: The  
MRI Jungle**

**Contrast Agents in MRI**

**Artefacts in MRI**

**MRI Advantages  
and Disadvantages**

**Take Home Messages**

**References**

**Test Your Knowledge**



**Hyperlinks**



**Core Knowledge**



**Further Knowledge**



**Attention**



**Compare**



**Questions**



**References**



# eBook for Undergraduate Education in Radiology

## Based on the ESR Curriculum for Undergraduate Radiological Education

Chapter: **Magnetic Resonance Imaging**

### Author

Bénédicte MA Delattre (1)  
Andrea Laghi (2)  
András Kincses (3,4)  
Minerva Becker (1)

### Affiliation

- (1) Division of Radiology, Diagnostic Department, Geneva University Hospitals, Geneva University, Geneva, Switzerland
- (2) Sapienza University, Rome, Italy
- (3) Department of Radiology, Albert Szent-Györgyi Faculty of Medicine and Clinical Center, University of Szeged, Szeged, Hungary
- (4) Institute of Biophysics, HUN-REN Biological Research Centre, Szeged, Hungary

benedicte.delattre@hug.ch  
andrea.laghi@uniroma1.it  
a.kincses08@gmail.com  
minerva.becker@hug.ch



## Chapter Outline

MR System

MR Principle

The Spin Echo Sequence

The Gradient  
Echo Sequence

Other Sequences: The  
MRI Jungle

Contrast Agents in MRI

Artefacts in MRI

MR Advantages  
and Disadvantages

Take Home Messages

References

Test Your Knowledge



## Chapter Outline

- **MRI System**
  - The Basics
  - Safety and Access Restriction
  - Components of the MRI System
- **MRI Principle**
  - Signal Production
  - Relaxation
  - Signal Reception
  - Spatial Encoding – Z Direction
  - Spatial Encoding – X and Y Directions
  - Relationship between Signal and Image
  - Gradients and MRI Noise
- **The Spin Echo Sequence**
  - What is an Echo?
  - Importance of TE & TR for Contrast?
  - MRI sequences: Why Are They so Long?
- **The Gradient Echo Sequence**
- **Other Sequences : The MRI Jungle**
  - Inversion Recovery (IR)
  - Diffusion Weighted Imaging (DWI)
  - Diffusion Tensor Imaging (DTI)
  - Time-of-Flight (ToF)
  - Functional MRI (fMRI)
  - Spectroscopy
- **Contrast Agents in MRI**
  - Perfusion Imaging (PI)
- **Artefacts in MRI**
- **MRI Advantages and Disadvantages**
- **Take-Home Messages**
- **References**
- **Test Your Knowledge**

## Chapter Outline

**MRI System**

**MRI Principle**

**The Spin Echo Sequence**

**The Gradient  
Echo Sequence**

**Other Sequences: The  
MRI Jungle**

**Contrast Agents in MRI**

**Artefacts in MRI**

**MRI Advantages  
and Disadvantages**

**Take Home Messages**

**References**

**Test Your Knowledge**




### Magnetic Resonance Imaging (MRI) System: The Basics



MRI is a non-invasive sophisticated technique that uses powerful magnetic fields to image the human body.

An MRI scanner is composed of 3 main parts :

- **Main magnet** => to produce the main static magnetic field ( $B_0$ ).
- **Gradient coils** => to produce deliberate variations in  $B_0$
- **Radiofrequency (RF) coils** => which act like the *antennas* of the MRI system : they transmit the RF field, and they receive the resulting signal.



#### Main magnet



The superconductive magnet (superconductive => no resistance to electricity) produces a high intensity magnetic field called " $B_0$ ". The magnet is cooled with liquid Helium (and liquid Nitrogen). It is used to generate a net magnetisation of tissue inside the bore. The bore size is 60-70 cm in diameter.



Constant electric current in a wire generates a static magnetic field (Biot-Savart law). The magnetic field strengths is proportional to the electric current.

Order of magnitude for magnetic field strengths:

- Earth magnetic field at latitude 0°: 31  $\mu$ T
- Fridge magnet : 5 mT
- Junkyard/scrap magnet : 1T
- Medical MRI : most often 1.5T and 3.0T, rarely 7.0T



Contraindications or restrictions for MRI :

- Claustrophobia
- Ferromagnetic metal in the body
- Some pacemakers, electronic implants, ...

### Chapter Outline

- ▶ **MRI System**
  - ▶ The Basics

**MRI Principle**

**The Spin Echo Sequence**

**The Gradient Echo Sequence**

**Other Sequences: The MRI Jungle**

**Contrast Agents in MRI**

**Artefacts in MRI**

**MRI Advantages and Disadvantages**

**Take Home Messages**



**References**

**Test Your Knowledge**





## MRI System: The Basics



The walls of the MRI magnet room (A) have layers which perform different functions: magnetic shielding to confine the stationary magnetic field, RF shielding to hinder electromagnetic noise to enter or exit the magnet room and acoustic shielding to restrict noise transmission beyond the magnet room. The control room (B) is located immediately outside the magnet room. It contains the operator console, computer equipment, communication devices, monitors (ECG and O<sub>2</sub>).

## Chapter Outline

- ▶ **MRI System**
  - ▶ The Basics
- MRI Principle**
- The Spin Echo Sequence**
- The Gradient Echo Sequence**
- Other Sequences: The MRI Jungle**
- Contrast Agents in MRI**
- Artefacts in MRI**
- MRI Advantages and Disadvantages**
- Take Home Messages**
- References**
- Test Your Knowledge**





## MRI System: Safety and Access Restriction



- The magnet is always ON !
- The main magnetic field  $B_0$  is always active. Never approach the field with a **ferromagnetic\*** object.
- The attraction force associated with the torque will pull the object through the main magnet with **uncontrollable** force: projectile effect or missile effect.
- **Past incidents unfortunately killed people !**
- This explains why safety rules around MRI are very strict!
- Patients undergoing MRI examinations must remove all metallic objects. Some radiology departments use ferromagnetic detection devices.

\* Ferromagnetic objects contain :  
- Iron, Cobalt, Nickel  
- Alloys of these components



MRI accident on a 1.5T MR system. A floor polishing machine was attracted by the magnetic field. It could only be removed by ramping down the magnetic field. Shown is the back side (head end) of the MRI. Reproduced from: [https://commons.wikimedia.org/wiki/File:MRI\\_accident\\_on\\_a\\_1.5\\_Tesla\\_MR\\_system.jpg](https://commons.wikimedia.org/wiki/File:MRI_accident_on_a_1.5_Tesla_MR_system.jpg)

## Chapter Outline

- ▶ **MRI System**
  - ▶ Safety and Access Restriction

### MRI Principle

### The Spin Echo Sequence

### The Gradient Echo Sequence

### Other Sequences: The MRI Jungle

### Contrast Agents in MRI

### Artefacts in MRI

### MRI Advantages and Disadvantages


### Take Home Messages

### References

### Test Your Knowledge




## MRI System: Components




### Radiofrequency coil

Produces also a varying magnetic field that is used to tip the net magnetisation perpendicularly to the main magnetic field  $B_0$ .



### Gradient coils

Produce varying magnetic fields in 3 spatial directions (x, y, z).  
Used to spatially encode the MRI signal



We'll see on the next pages how each of these parts contributes to the production of images =>

## Chapter Outline

- ▶ **MRI System**
  - ▶ Components
- MRI Principle**
- The Spin Echo Sequence**
- The Gradient Echo Sequence**
- Other Sequences: The MRI Jungle**
- Contrast Agents in MRI**
- Artefacts in MRI**
- MRI Advantages and Disadvantages**
- Take Home Messages**
- References**
- Test Your Knowledge**



## MRI Principle





The nucleus of an atom is composed of protons (positive charge) and neutrons (no charge) which all rotate around their own axis. The electrons (negative charge) revolve around the nucleus, and they also rotate around their own axis.

The rotation of all these particles produces an angular momentum of rotation, which is called **spin**. A spin is a fundamental property of atoms like mass or electrical charge. Spin comes in multiples of  $\frac{1}{2}$ .

As the proton has a positive charge and as it rotates continuously, it creates a small magnetic field, called **magnetic moment** (i.e., it behaves like a tiny magnet with a north and south pole).

- There is a natural abundance of  $\text{H}_2\text{O}$  in biological tissues and, therefore, an abundance of  $\text{H}^1$ .
- $\text{H}^1$  **mainly occurs** in water in the human body.
- Human body composition  $\Rightarrow$  ~ 60% - 70% water (2  $\text{H}^1$ )
- $\text{H}^1$  has a **large** magnetic moment.
- The magnetic property of  $\text{H}^1$  is used to mainly image the water distribution of tissues in the body with MRI.



## Chapter Outline

MRI System

► MRI Principle

The Spin Echo Sequence

The Gradient  
Echo Sequence

Other Sequences: The  
MRI Jungle

Contrast Agents in MRI

Artefacts in MRI

MRI Advantages  
and Disadvantages

Take Home Messages

References

Test Your Knowledge






## MRI Principle



When biological tissue is placed in a strong magnetic field, a **net magnetization vector** is created. To effectively explain this phenomenon, quantum mechanics is required, which is beyond the scope of this chapter.


This effect applies to atoms with **specific magnetic moment properties** i.e., nucleus with spin quantum number =  $\frac{1}{2}$  :  $H^1$  /  $C^{13}$  /  $N^{15}$  /  $O^{17}$  /  $Na^{23}$  / ...

Alignment **parallel**  $\uparrow$  or **anti-parallel**  $\downarrow$  to the magnetic field, corresponds to two different energy states. Most protons align parallel to  $B_0$  as this requires less energy than the antiparallel alignment. The net magnetization  $\vec{M}$  is created by the fraction of spin in excess in one of the energy states :



Outside the magnetic field

$B_0$



Inside the magnetic field

$\vec{M}$



EXAMPLE : fraction in excess  
@ 3T  $\sim 10$  on  $10^6$  !

## Chapter Outline

MRI System

► MRI Principle

The Spin Echo Sequence

The Gradient Echo Sequence

Other Sequences: The MRI Jungle

Contrast Agents in MRI

Artefacts in MRI

MRI Advantages and Disadvantages

Take Home Messages

References

Test Your Knowledge




## Signal Production



### Excitation ...

- To create a signal from the tissue, a radiofrequency (RF) wave is used. It is tuned to the resonance frequency of the spins called « Larmor frequency »  $f$ , defined by :
$$f = gB_0$$
- Where  $g$  is the gyromagnetic ratio ( $g = 42.58 \text{ MHz/T}$ ) and  $B_0$  the magnetic field strength.
  - At 1.5T,  $f = 64 \text{ MHz}$
  - At 3T,  $f = 128 \text{ MHz}$



### ... and relaxation

## Chapter Outline

### MRI System

### ► MRI Principle ► Signal Production

### The Spin Echo Sequence

### The Gradient Echo Sequence

### Other Sequences: The MRI Jungle

### Contrast Agents in MRI

### Artefacts in MRI

### MRI Advantages and Disadvantages

### Take Home Messages

### References

### Test Your Knowledge




### Relaxation

Relaxation is happening by two simultaneous but distinct processes.



#### T1 relaxation

Spin energy is dispersed into its environment (mainly nucleus and other atoms), the magnetization is recovering its initial state along  $B_0$  (longitudinal magnetization). The mechanism by which  $M_z$  exponentially relaxes from a higher energy state to thermodynamic equilibrium is also called **spin-lattice relaxation**.



#### T2 relaxation

Magnetization flipped in the transverse plane is reduced due to spin dephasing. Phase coherence is lost, reducing net magnetization in the x-y plane (while net magnetization is re-growing in the z direction through T1 relaxation !). The mechanism by which  $M_{xy}$  exponentially decays towards its equilibrium value is also called **spin-spin relaxation**.



### Chapter Outline

MRI System

► MRI Principle  
    ► Relaxation

The Spin Echo Sequence

The Gradient Echo Sequence

Other Sequences: The MRI Jungle

Contrast Agents in MRI

Artefacts in MRI

MRI Advantages and Disadvantages

Take Home Messages

References

Test Your Knowledge





## Relaxation



T1 is the the time constant for regrowth of  $M_z$  (longitudinal magnetization)

T2 is the time constant for decay/dephasing of  $M_{x,y}$  (transverse magnetization)

T1 and T2 relaxation times depend on the environment, they are characteristic for different tissues !

Below some examples of T1 and T2 relaxation values at 1.5T

At a main field of 1.5 T

Tissue type	Approximate T <sub>1</sub> value in ms	Approximate T <sub>2</sub> value in ms
Adipose tissues	240-250	60-80
Whole blood (deoxygenated)	1350	50
Whole blood (oxygenated)	1350	200
Cerebrospinal fluid (similar to pure water)	4200 - 4500	2100-2300
Gray matter of cerebrum	920	100
White matter of cerebrum	780	90
Liver	490	40
Kidneys	650	60-75
Muscles	860-900	50

From: [https://en.wikipedia.org/wiki/Relaxation\\_\(NMR\)](https://en.wikipedia.org/wiki/Relaxation_(NMR))

## Chapter Outline

MRI System

► MRI Principle  
    ► Relaxation

The Spin Echo Sequence

The Gradient  
Echo Sequence

Other Sequences: The  
MRI Jungle

Contrast Agents in MRI

Artefacts in MRI

MRI Advantages  
and Disadvantages

Take Home Messages

References


Test Your Knowledge



## Signal Reception



The periodic signal accompanying the relaxation of the excited net magnetization can be recorded by a coil.



Coil = loop of conductive wire



$M_{xy}$  = Signal measured

Inductive current created by time varying magnetization

## Chapter Outline

### MRI System

#### ► MRI Principle ► Signal Reception

### The Spin Echo Sequence

### The Gradient Echo Sequence

### Other Sequences: The MRI Jungle

### Contrast Agents in MRI

### Artefacts in MRI

### MRI Advantages and Disadvantages

### Take Home Messages

### References

### Test Your Knowledge



### Signal Reception



- Dedicated coils are used for each application :
- Modern receiving coils contain several small coils (also called channels), each one receiving the emitted signal. Such configurations help to achieve high signal to noise ratio, as well as a large coverage of the anatomy to investigate.


Head coil (32 channels)



Spine coil (usually included in the patient table)



Hand/wrist coil



Shoulder coil (16 channels)



Foot/ankle coil



### Chapter Outline

#### MRI System

- ▶ **MRI Principle**
  - ▶ Signal Reception

#### The Spin Echo Sequence

#### The Gradient Echo Sequence

#### Other Sequences: The MRI Jungle

#### Contrast Agents in MRI

#### Artefacts in MRI

#### MRI Advantages and Disadvantages

#### Take Home Messages

#### References

#### Test Your Knowledge






## Spatial Encoding – Z Direction





- At this stage, signal is provided by the whole volume of tissue excited by the RF coil.
- Remember : the system is composed of 3 gradient coils, one for each geometrical dimension (x, y, z).
- The gradient coils are used to add some spatial encoding to the signal !
- How does it work ?

Example with slice encoding

Volume excited by the RF tuned at  $\omega_0$  (without any gradient)




Addition of magnetic field varying in z direction with the z-gradient



$$\omega_1 = \omega_0 + \omega_G(z)$$

$\omega_2$




$\omega_3$

$\omega_4$

RF excitation @  $\omega_2$   
Results in slice selection

To spatially select signal coming from one slice, we tune the RF to the corresponding modified frequency, here  $\omega_2$  for example.



## Chapter Outline

**MRI System**

► **MRI Principle**  
    ► Spatial Encoding  
       Z Direction

**The Spin Echo Sequence**

**The Gradient Echo Sequence**

**Other Sequences: The MRI Jungle**

**Contrast Agents in MRI**

**Artefacts in MRI**

**MRI Advantages and Disadvantages**

**Take Home Messages**

**References**


**Test Your Knowledge**



## Spatial Encoding – X and Y Directions



To add spatial information on the two other dimensions, x- and y-gradients are also used at specific timing before and during signal reception. They are used to add **spatially varying dephasing** (in the so-called **phase encoding direction**) and spatially varying frequency (in the so-called **frequency direction**).



## Chapter Outline

### MRI System

- ▶ **MRI Principle**
  - ▶ Spatial Encoding X and Y Directions

### The Spin Echo Sequence

### The Gradient Echo Sequence

### Other Sequences: The MRI Jungle

### Contrast Agents in MRI

### Artefacts in MRI

### MRI Advantages and Disadvantages

### Take Home Messages

### References

### Test Your Knowledge



## Relationship between Signal and Image




The measured signal is in the frequency space, so called "k-space". The "k" stands for a number that keeps the gradient spatial encoding information. This "k-space" can be translated into the final image using the **Fourier transform**.



**REMEMBER:** Until now we have measured magnetization of spins with varying additional gradients, to create information contained in the frequency space (k-space). The image will be created by the Fourier transform of this k-space.


K-space



Fourier  
transform



image



## Chapter Outline

### MR System

- ▶ **MR Principle**
  - ▶ Relationship between Signal and Image

### The Spin Echo Sequence

### The Gradient Echo Sequence

### Other Sequences: The MRI Jungle

### Contrast Agents in MRI

### Artefacts in MRI

### MR Advantages and Disadvantages

### Take Home Messages


### References

### Test Your Knowledge






### Gradients and MRI Noise



Gradient coils used to add spatial information to the signal produce varying magnetic fields during image acquisition. These variations make the gradient coils vibrate. This is the origin of the loud noise heard during MR image acquisition.

The patient must wear hearing protection during the exam!



140 dB = Airplane taking off  
130 dB = MRI  
110 dB = Concert or nightclub  
95 dB = school cafeteria  
85 dB = lawn mower  
80 dB = car  
60 dB = conversation

### Chapter Outline

#### MRI System

- ▶ **MRI Principle**
  - ▶ Gradients and MRI Noise

#### The Spin Echo Sequence

#### The Gradient Echo Sequence

#### Other Sequences: The MRI Jungle

#### Contrast Agents in MRI

#### Artefacts in MRI

#### MRI Advantages and Disadvantages

#### Take Home Messages

#### References

#### Test Your Knowledge




## The Spin Echo (SE) Sequence



### What is an Echo?

When the magnetization is coming back to equilibrium after the application of an RF pulse, the signal produced is called a « free induction decay ».

When a second RF pulse is used, in particular a  $180^\circ$  pulse, or **refocusing pulse**, then an echo is created ! The time between the first RF pulse and the echo is called the **echo time**.



## Chapter Outline

[MRI System](#)

[MRI Principle](#)

► [The Spin Echo Sequence](#)

► [What is an Echo?](#)

[The Gradient Echo Sequence](#)

[Other Sequences: The MRI Jungle](#)

[Contrast Agents in MRI](#)

[Artefacts in MRI](#)

[MRI Advantages and Disadvantages](#)

[Take Home Messages](#)

[References](#)

[Test Your Knowledge](#)



## The Spin Echo (SE) Sequence

Between the  $90^\circ$  and the  $180^\circ$  pulses spins are dephasing. Fast spinning components dephase more than slow spinning components. The  $180^\circ$  pulse rephases the spins by "reversing" the dephasing, so the fast-spinning components are then regaining phase to join the slow-spinning components.

Analogy => The Hare and the Tortoise

The Hare : fast spinning component

The Tortoise : slow spinning component




Illustration from:  
[https://en.wikipedia.org/wiki/Spin\\_echo](https://en.wikipedia.org/wiki/Spin_echo)

[Click to Play Video in Browser \(External\)](#)



[Click to Play Video in Browser \(External\)](#)



## Chapter Outline

[MRI System](#)

[MRI Principle](#)

▶ [The Spin Echo Sequence](#)  
▶ What is an Echo?

[The Gradient Echo Sequence](#)

[Other Sequences: The MRI Jungle](#)

[Contrast Agents in MRI](#)

[Artefacts in MRI](#)

[MRI Advantages and Disadvantages](#)

[Take Home Messages](#)

[References](#)

[Test Your Knowledge](#)






### The Spin Echo (SE) Sequence

As explained previously, the signal coming from the echo is emitted by the whole excited volume! Gradients are added to encode the spatial origin of the signal. The timing diagram of when and where these gradients are applied regarding the excitation RF pulses and the echo reading represents the "MR sequence".

There are 2 important timing parameters in the sequence :

- Time between the excitation pulse and the echo : Echo time, TE
- Time between two successive excitation pulses : Repetition time, TR.
- TE and TR are chosen by the operator!



### Chapter Outline

**MRI System**

**MRI Principle**

► **The Spin Echo Sequence**  
► What is an Echo?

**The Gradient Echo Sequence**

**Other Sequences: The MRI Jungle**

**Contrast Agents in MRI**

**Artefacts in MRI**

**MRI Advantages and Disadvantages**

**Take Home Messages**

**References**

**Test Your Knowledge**



## The Spin Echo (SE) Sequence


### Importance of TE and TR for Contrast

Long TE contributes to T2 contrast.

- Signal is less prominent for fast dephasing spins (low T2) than for slow dephasing spins (high T2).
- Short TE does not allow spins to dephase, no contribution of T2 contrast in the signal.

Short TR contributes to T1 contrast.

- Magnetization regrowing is not complete, slowly growing magnetization will give less signal (high T1) than fast growing magnetization (low T1).
- Long TR lets the longitudinal magnetization regrowth to its original state, no contribution to T1 contrast.



## Chapter Outline

MRI System

MRI Principle

► **The Spin Echo Sequence**  
    ► Importance of TE and TR for Contrast

The Gradient Echo Sequence

Other Sequences: The MRI Jungle

Contrast Agents in MRI

Artefacts in MRI

MRI Advantages and Disadvantages

Take Home Messages

References


Test Your Knowledge



## The Spin Echo (SE) Sequence


### Importance of TE and TR for Contrast

To obtain an image without T1 nor T2 contrast but only sensitive to proton density, a short TE and a long TR should be used.



Summary:

- Short TE, short TR : T1-weighted image
- Long TE, long TR : T2-weighted image
- Short TE, Long TR : Proton density weighted image
- (Long TE, short TR is not used in practice)



## Chapter Outline

[MRI System](#)

[MRI Principle](#)

▶ [The Spin Echo Sequence](#)  
▶ Importance of TE and TR for Contrast

[The Gradient Echo Sequence](#)

[Other Sequences: The MRI Jungle](#)

[Contrast Agents in MRI](#)

[Artefacts in MRI](#)

[MRI Advantages and Disadvantages](#)

[Take Home Messages](#)


[References](#)

[Test Your Knowledge](#)





## The Spin Echo (SE) Sequence



## Chapter Outline

MRI System

MRI Principle

► **The Spin Echo Sequence**  
► Importance of TE and TR for Contrast

**The Gradient Echo Sequence**

**Other Sequences: The MRI Jungle**

**Contrast Agents in MRI**

**Artefacts in MRI**

**MRI Advantages and Disadvantages**

**Take Home Messages**

**References**

**Test Your Knowledge**



## MRI Sequences: Why are They so Long?



The **acquisition time** of an MRI sequence depends mainly on TR, on the number of phase encoding lines (matrix) and on the number of slices :

$$\text{Acquisition time} = \text{TR} \cdot \text{NPy} \cdot \text{Nslices}$$

- TR = Repetition time
- NPy = Number of phase encoding lines
- Nslices = Number of slices

Example for a T1 sequence : TR = 500 ms, 128 matrix size, 10 slices :  
Acquisition time = 0.5 sec \* 128 \* 10 = 10 min 40 sec !

MRI is a slow acquisition technique !

In practice, several techniques have been developed to accelerate sequence acquisition :  
Typical acquisition time for a 2D sequence covering the whole brain is 2 to 4 minutes.  
Acquisition for 3D sequences is longer, typically around 4 to 6 minutes.

## Chapter Outline

MRI System

MRI Principle

► **The Spin Echo Sequence**  
    ► MRI Sequences: Why are They so Long?

The Gradient Echo Sequence

Other Sequences: The MRI Jungle

Contrast Agents in MRI

Artefacts in MRI

MRI Advantages and Disadvantages


Take Home Messages

References

Test Your Knowledge



## The Gradient Echo (GRE) Sequence



The Gradient Echo (GRE) sequence doesn't use a  $180^\circ$  RF pulse to refocus the dephasing spins but uses instead a gradient to dephase and then rephase the spins, thus creating an echo.


This can shorten a lot the TE and TR and make faster images !

But... it adds dephasing errors, so it is more prone to artifacts.

Phase encoding :  
Gradient in phase direction applied to encode y direction, this gradient varies at each « TR » to encode the different lines in phase direction.

Slice selection :  
Gradient in slice direction z is applied during excitation pulse

Rephasing gradient in slice direction



Frequency encoding :  
Gradient in frequency direction is applied to encode x direction, this gradient is applied at the same time as the echo is produced and measured.

Dephasing in frequency direction, gradient is applied to compensate the rephasing produced during the echo.

## Chapter Outline

MRI System

MRI Principle

The Spin Echo Sequence

► The Gradient Echo Sequence

Other Sequences: The MRI Jungle

Contrast Agents in MRI

Artefacts in MRI

MRI Advantages and Disadvantages

Take Home Messages

References

Test Your Knowledge






## The Gradient Echo (GRE) Sequence




Since there is no more refocusing RF pulse, spin dephasing is now due to local magnetic field inhomogeneities in addition to the T2 effect. Therefore, the signal decreases with the T2\* constant. The difference between T2 and T2\* is that T2 is the ideal spin-spin relaxation caused by atomic/molecular interactions whereas T2\* is the observed T2 (i.e., T2 affected by local field inhomogeneities



This effect allows a high sensitivity to local magnetic inhomogeneities typically around blood degradation products and calcifications that are locally disturbing the magnetic field.



T2\* is always  $\leq$  T2 !



## Chapter Outline

[MRI System](#)

[MRI Principle](#)

[The Spin Echo Sequence](#)

► [The Gradient Echo Sequence](#)

[Other Sequences: The MRI Jungle](#)

[Contrast Agents in MRI](#)

[Artefacts in MRI](#)

[MRI Advantages and Disadvantages](#)

[Take Home Messages](#)

[References](#)


[Test Your Knowledge](#)



## Other Sequences : The MRI Jungle



Many different types of sequences are used for MR imaging named differently by each vendor.  
Nearly all are derived from SE or GRE sequences.



## Chapter Outline

MRI System

MRI Principle

The Spin Echo Sequence

The Gradient  
Echo Sequence

► Other Sequences: The  
MRI Jungle

Contrast Agents in MRI

Artefacts in MRI

MRI Advantages  
and Disadvantages

Take Home Messages

References


Test Your Knowledge




## Inversion Recovery (IR)



Adding a preparation pulse **before** the acquisition of the signal can increase tissue contrast or remove signal from a specific tissue. The IR-sequence uses an inversion pulse ( $180^\circ$ ) before the sequence to invert the entire magnetization. IR techniques are widely used in neuroradiology, head and neck and cardiac Imaging applications.



IR-prepared 3D T1-weighted GRE sequence: Inversion pulse increases white matter/grey matter contrast (this image is acquired after contrast agent injection).



Fluid Attenuated IR (FLAIR) : Inversion pulse is used to remove signal from cerebrospinal fluid. Hyperintense signal in white matter lesions is more visible.



STIR : Inversion pulse removes signal from fat. Hyperintense signal due to fluids or oedema is more visible.

## Chapter Outline

**MRI System**

**MRI Principle**

**The Spin Echo Sequence**

**The Gradient Echo Sequence**

▶ **Other Sequences: The MRI Jungle**  
▶ Inversion Recovery

**Contrast Agents in MRI**

**Artefacts in MRI**

**MRI Advantages and Disadvantages**

**Take Home Messages**

**References**

**Test Your Knowledge**





## Diffusion Weighted Imaging (DWI)

Diffusion is defined as the transport of matter resulting from the migration of **atoms** due to the **random movements** caused by differences in temperature or concentration.

MRI can be sensitive to water molecule diffusion, which depends on the environment (intracellular, extracellular, intravascular water). The MRI signal can, therefore, reflect cellular membrane integrity or cellular density.



One of the successes of MRI is the ability to detect cellular oedema in the very early stages of a stroke, before any other type of imaging modality can show the stroke.



$\leq$  Diffusion-weighted Images (b1000) in a normal brain (A) and in a patient with **stroke in the middle cerebral artery territory** (B).

In stroke, due to hypoperfusion, there is failure of the sodium/potassium pumps of cell membranes in the affected areas. This creates an influx of intracellular water, thus decreasing diffusion movement ("restricted diffusion"). On MRI, there is an increased signal on the DWI image.



Hypercellular tumours also show restricted diffusion because the free movement of water molecules is hindered by the densely packed cells.



## Chapter Outline

**MRI System**

**MRI Principle**

**The Spin Echo Sequence**

**The Gradient Echo Sequence**

► **Other Sequences: The MRI Jungle**  
    ► Diffusion Weighted Imaging

**Contrast Agents in MRI**

**Artefacts in MRI**

**MRI Advantages and Disadvantages**

**Take Home Messages**

**References**

**Test Your Knowledge**





### Diffusion Tensor Imaging (DTI)




Biological tissues are highly **anisotropic** => i.e., the diffusion rate is NOT the same in all directions.

Water diffusion in the brain is **constrained by fibres**, The MRI signal is sensitive to the **preferential direction of motion** of water molecules. This can be used to «track» fibres and depict white matter tracts. The anatomical orientation of axons and fibres is coded with colours on DTI images, each colour corresponding to a specific direction of the fibres:

- **Red** => transverse orientation
- **Green** => anterior-posterior or posterior-anterior
- **Blue** => craniocaudal orientation

Fibre tracts are then reconstructed depending on the clinical question using a dedicated software. Quantitative measures can be obtained, e.g., measuring the **fractional anisotropy (FA)** which is thought to reflect fibre density, myelination and axonal diameter.



Example of a DTI examination in trigeminal neuralgia. DTI images with overlaid colour-by-orientation fibres at the mid-pontine level (A). Reconstructed tracts of the **trigeminal nerves** onto colour-by-code orientation (B).

## Chapter Outline

**MRI System**

**MRI Principle**

**The Spin Echo Sequence**

**The Gradient Echo Sequence**

► **Other Sequences: The MRI Jungle**

► Diffusion Tensor Imaging

**Contrast Agents in MRI**

**Artefacts in MRI**

**MRI Advantages and Disadvantages**

**Take Home Messages**

**References**


**Test Your Knowledge**



### Diffusion Tensor Imaging (DTI)



In addition to showing the 3D representation of fibre tracts, DTI can detect micro-structural changes in the absence of morphologic changes. It can reveal altered white matter connectivity and allows quantitative evaluation of the integrity of different brain circuits in a variety of conditions including tumours, demyelinating diseases, trauma, Parkinson disease, pain syndromes, depression and anxiety disorders and many more.



Example of a left temporo-insular low-grade glioma (LGG) and fibre tract involvement. 3D reconstruction of the tumour, which involves the fronto-occipital longitudinal fasciculus.

Reproduced from: Ius T et al. Risk Assessment by Pre-surgical Tractography in Left Hemisphere Low-Grade Gliomas. Front Neurol. 2021 Feb 15;12:648432. doi: 10.3389/fneur.2021.648432.




Image illustrating white matter abnormalities in adolescents with generalised anxiety disorder (GAD). Voxels are overlaid on the white matter skeleton (green). The regions of significant FA reduction in comparison to adolescents without GAD are shown in red.

Reproduced from: Liao, M. et al. White matter abnormalities in adolescents with generalized anxiety disorder: a diffusion tensor imaging study. BMC Psychiatry 14, 41 (2014). <https://doi.org/10.1186/1471-244X-14-41>

## Chapter Outline

**MRI System**

**MRI Principle**

**The Spin Echo Sequence**

**The Gradient Echo Sequence**

► **Other Sequences: The MRI Jungle**

► Diffusion Tensor Imaging

**Contrast Agents in MRI**

**Artefacts in MRI**

**MRI Advantages and Disadvantages**

**Take Home Messages**

**References**

**Test Your Knowledge**




## Time of Flight (TOF) MR Angiography (MRA)




The Time of Flight (TOF) MRA sequence allows visualization of flowing blood in vessels thus providing angiographic images **without** the need of injecting contrast agents. The TOF sequence is based on the principle of **flow related enhancement** (i.e., fresh blood has a high initial magnetization as opposed to stationary tissues, which are magnetically saturated by multiple repetitive RF pulses). On the TOF sequence, the signal of inflowing blood appears very bright (see below). The maximal flow enhancement occurs when the vessel is perpendicular to the imaging plane.



TOF is one of the most useful techniques for non-contrast neurovascular and peripheral MRA => See e-book chapter on vascular imaging



TOF image : Maximum Intensity Projection (MIP) of the polygon of Willis: axial (A), coronal (B) and sagittal (C) views.



## Chapter Outline

**MRI System**

**MRI Principle**

**The Spin Echo Sequence**

**The Gradient Echo Sequence**

► **Other Sequences: The MRI Jungle**

► Time of Flight MR Angiography

**Contrast Agents in MRI**

**Artefacts in MRI**

**MRI Advantages and Disadvantages**

**Take Home Messages**

**References**

**Test Your Knowledge**





### Functional MRI (fMRI)




Blood Oxygenation Dependent (BOLD) imaging, is the standard **functional** MRI modality, which provides information about cerebral areas that are activated while performing certain tasks. For example, it is possible to identify the areas of language in the brain. This is very useful to determine if an area is impacted by surgery or if a lesion is located in immediate vicinity of an area that needs to be resected.


BOLD imaging is based on the principle that if a task leads to an increase in the activity of a specific brain region, there is an **initial drop in oxygenated haemoglobin and an increase in CO<sub>2</sub> and deoxygenated haemoglobin**. After a delay of a few seconds, the increased cerebral blood flow (CBF) delivers a **surplus** of oxygenated haemoglobin, which “washes away” deoxyhaemoglobin.

Oxygenated and deoxygenated haemoglobin differ significantly with respect to their **paramagnetic** properties.

T2\* sequences are used to detect these differences, which are in the range of 1-5%.



Example of BOLD fMRI maps obtained in a patient with a **high-grade glioma** and silent word generation task producing activation of the left prefrontal cortex and Broca's area. Figure courtesy José Manuel Baiao Boto, Division of Neuroradiology, Geneva University Hospitals.



Example of BOLD fMRI maps obtained in a patient with a **high-grade glioma** and right-sided finger tapping. The contralateral (left) sensorimotor cortex is most strongly activated. Figure courtesy José Manuel Baiao Boto, Division of Neuroradiology, Geneva University Hospitals.

### Chapter Outline

**MRI System**

**MRI Principle**

**The Spin Echo Sequence**

**The Gradient Echo Sequence**

▶ **Other Sequences: The MRI Jungle**

▶ Functional MR

**Contrast Agents in MRI**

**Artefacts in MRI**

**MRI Advantages and Disadvantages**

**Take Home Messages**

**References**

**Test Your Knowledge**






### MR Spectroscopy (MRS)




MR Spectroscopy (MRS) is a method to measure the chemical composition of tissue. It allows measurement of metabolites **in vivo** in **specific brain regions** such as N-acetyl aspartate (NAA), Choline (Cho), Creatine (Cr), and others. MRS uses the fact that the proton resonant frequency is **slightly different** for each metabolite compared to water.

MRS is **mostly used in the brain**, but it is not restricted to this area. Advances were made to increase the spatial resolution and even create metabolite maps of the brain. Most common indications for MRS include imaging of gliomas, post-radiation changes, ischemia, white matter and mitochondrial diseases. MRS increases specificity and correlates with the histologic grade of a tumour.



MRS obtained in a patient with a **high-grade glioma in the right basal ganglia** (A) showing metabolite changes. As the tumour grade increases, NAA and Cho decrease whereas lipids (Lip) and lactate (Lac) increase. **Normal MRS metabolites in the left basal ganglia** (B). Measurements on the left are used as control. Note that the normal Cho, Cr and NAA peaks are on a line which has a 45-degree angle with the x- axis (Hunter's angle). Figure courtesy José Manuel Baiao Boto, Division of Neuroradiology, Geneva University Hospitals.



### Chapter Outline

**MRI System**

**MRI Principle**

**The Spin Echo Sequence**

**The Gradient Echo Sequence**

► **Other Sequences: The MRI Jungle**

► MR Spectroscopy

**Contrast Agents in MRI**

**Artefacts in MRI**

**MRI Advantages and Disadvantages**

**Take Home Messages**

**References**

**Test Your Knowledge**




### MRI Contrast Agents




Contrast agents (CA)s used in MRI are mostly based on Gadolinium chelates. Gadolinium is paramagnetic and has the property of reducing T1 relaxation of surrounding tissues, thus rendering them hyperintense on T1 contrast. At high concentrations, Gadolinium-based CA also shortens T2 relaxation time. It normally stays extracellular in the circulation and microcirculation system and it is excreted by the kidneys.

T1 SE



Enhancing signal in the lesion after contrast injection



T1 SE with Gd



#### Safety : Nephrogenic Systemic Fibrosis (NSF)

In 2006 Gadolinium-based contrast agents were recognized to be the potential triggers of a late inflammatory and fibrotic disease of soft tissues in patients with severe renal function impairment : nephrogenic systemic fibrosis (NSF). Even though NSF **is rare** it remains **mandatory to screen patients for renal dysfunction** prior to Gd-chelate administration, and to assess risk and benefits prior to contrast agent injection. => See e book chapter on contrast media.

## Chapter Outline

MRI System

MRI Principle

The Spin Echo Sequence

The Gradient Echo Sequence

Other Sequences: The MRI Jungle

► Contrast Agents in MRI

Artefacts in MRI

MRI Advantages and Disadvantages

Take Home Messages

References

Test Your Knowledge



## MRI Contrast Agents



### Gadolinium (Gd) Accumulation in the Central Nervous System

Accumulation of Gd in central nervous system (CNS), basically in the basal ganglia, was reported in patients with multiple administrations of Gd-chelates (2014).

The trans-metalation reaction is a possible mechanism by which the Gd ion is extracted from the chelate by another cation. It has been shown that Gd-chelates with a linear configuration are more at risk to accumulate in the CNS than macrocyclic chelates => see e book chapter on contrast media.

This is the reason why the European Medicines Agency recommended to suspend or limit the use of commercially available linear Gd-based contrast agents.

Even though Gd accumulation is now well described, there is **no evidence of clinical short- or long-term effects**. Precaution principle has to be applied by reducing amount and frequency of Gd injection when possible.



Kanda T, Ishii K, Kawaguchi H, et al. High signal intensity in the dentate nucleus and globus pallidus on unenhanced T1-weighted MR images: relationship with increasing cumulative dose of a gadolinium-based contrast material. Radiology 2014; 270:834-841. (Landmark first report of Gd accumulation in the brain).

## Chapter Outline

MRI System

MRI Principle

The Spin Echo Sequence

The Gradient Echo Sequence

Other Sequences: The MRI Jungle

► Contrast Agents in MRI

Artefacts in MRI

MRI Advantages and Disadvantages

Take Home Messages

References

Test Your Knowledge






### MRI Perfusion Weighted Imaging (PWI)



MRI PWI encompasses different MRI techniques used to assess the perfusion of tissues by blood. To assess perfusion, contrast-enhanced techniques and non-contrast enhanced techniques (e.g., arterial spin labelling, ASL) can be applied.

**Dynamic Susceptibility Contrast (DSC)** MRI PWI relies on the signal loss induced by a bolus of Gd-based contrast agent on T2\*-weighted sequences. The calculated parameters include a **Time signal Intensity Curve (TIC)** from which **cerebral blood volume (CBV** = volume of blood in a given brain tissue amount in ml blood/100g brain tissue), **cerebral blood flow (CBF** = CBV per unit of time, in ml blood/100g brain tissue/minute) and other parameters are calculated. These parameters are then used to create colour maps of regions of interest. Due to the difficulty to precisely calculate CBV and CBF, most often CBV / CBF relative to an internal control, e.g., contralateral normal white matter are calculated (rCBV and rCBF). rCBV and rCBF have no units as they correspond to ratios.



T2\*-weighted DSC MRI perfusion in a patient with a glioblastoma. A. TICs obtained in different regions of interest (ROIs). B. rCBV colour map and the corresponding axial contrast enhanced T1 weighted image showing increased tumour perfusion. Figure courtesy José Manuel Baiao Boto, Division of Neuroradiology, Geneva University Hospitals.

### Chapter Outline

**MRI System**

**MRI Principle**

**The Spin Echo Sequence**

**The Gradient Echo Sequence**

**Other Sequences: The MRI Jungle**

► **Contrast Agents in MRI**  
► Perfusion Weighted Imaging

**Artefacts in MRI**

**MRI Advantages and Disadvantages**

**Take Home Messages**

**References**

**Test Your Knowledge**








## MRI Perfusion Weighted Imaging (PWI)



Dynamic Contrast Enhanced (DCE) MRI PWI is one of the most important MRI PWI techniques. Perfusion parameters are calculated on the basis of **T1 shortening effects** due to the bolus of Gd-based contrast agent passing through tissue. The following parameters are calculated: TICs, k-trans (= volume transfer constant from blood plasma to extravascular extracellular space), fractional volume of extravascular-extracellular space, and others. TICs are very useful for the characterisation of certain tumours. For example, certain TIC types can be found only in malignant tumours whereas other TIC types only in benign lesions.

DCE MRI PWI is mainly used for oncologic imaging.



T1-weighted Dynamic Contrast Enhanced (DCE) PWI in a patient with a diffusely infiltrating left parotid tumour. A. Time resolved dynamic sequence. B. ROIs placed for measurements (carotid artery, parotid tumour, muscle). C. Time-intensity curves (TICs) in the different regions of interest shown in B. TIC colours correspond to the ROIs indicated in B.

## Chapter Outline

MRI System

MRI Principle

The Spin Echo Sequence

The Gradient Echo Sequence

Other Sequences: The MRI Jungle

► Contrast Agents in MRI  
► Perfusion Weighted Imaging

Artefacts in MRI

MRI Advantages and Disadvantages

Take Home Messages

References

Test Your Knowledge



### Artefacts in MRI



The image acquisition process can be responsible for different artifacts in the image, some can easily be addressed while other not. MRI sequence optimization requires understanding of numerous parameters, this ability is essential to obtain images with the best mitigation of artefacts.

Recognition of these artefacts in the images is an important part of the radiologist experience !

Origin of artefacts can be separated in three categories :

- Technique :
  - Type of sequence
  - Parameters
- Patient :
  - Motion (uncontrollable),
  - Breathing, blood flow,
  - Implants, tattoo, piercing, ...
- Hardware :
  - Receiver coil, RF coil, gradient coils,
  - Faraday cage

Can be corrected

Can be mitigated with specific techniques

Has to be repaired

### Chapter Outline

**MRI System**

**MRI Principle**

**The Spin Echo Sequence**

**The Gradient  
Echo Sequence**

**Other Sequences: The  
MRI Jungle**

**Contrast Agents in MRI**

▶ **Artefacts in MRI**

**MRI Advantages  
and Disadvantages**

**Take Home Messages**

**References**

**Test Your Knowledge**



## Artefacts: Examples




Artefact due the technique : wrong parameters of the sequence.



Fold over artefact: nose (outside of the field of view) is projected in the centre of the image !

Artefact due to the patient : blood flow in arteries.



Flow artefact: signal from blood flowing in arteries is propagated in the phase encoding direction.

Artefact due to the patient : presence of braces



Susceptibility artefact: signal loss due to presence of metal in the mouth (braces), magnetic field perturbation extends largely outside the mouth.

## Chapter Outline

**MRI System**

**MRI Principle**

**The Spin Echo Sequence**

**The Gradient Echo Sequence**

**Other Sequences: The MRI Jungle**

**Contrast Agents in MRI**

► **Artefacts in MRI**

**MRI Advantages and Disadvantages**

**Take Home Messages**

**References**

**Test Your Knowledge**



## MRI Advantages and Disadvantages



- Non ionising modality suitable for follow-up examinations
- Excellent soft tissue contrast (ligaments, tendons, muscles, brain grey and white matter, ...)
- Different type of contrast images available (sensitive to fluid, with fat suppression, ...)
- Good image resolution, 2D images in any orientation and 3D images possible
- Anatomical but also functional imaging possible (diffusion, perfusion, fMRI, MRS, ...)



- Not all implants are allowed in the magnetic field
- Not suitable for claustrophobic patients (larger bore available nowadays)
- Noisy and generally long examinations
- More expensive than CT or X-ray
- Requires good knowledge of the technique (sequence optimization and artefact mitigation)

## Chapter Outline

MRI System

MRI Principle

The Spin Echo Sequence

The Gradient Echo Sequence

Other Sequences: The MRI Jungle

Contrast Agents in MRI

Artefacts in MRI

► MRI Advantages and Disadvantages

Take Home Messages

References

Test Your Knowledge





## Take-Home Messages



- MRI is a non-ionizing, non-invasive imaging modality.
- It provides excellent soft tissue contrast and offers unique anatomical and functional information.
- Some restrictions or contraindications exist for patients with implanted material or devices.
- The main magnet is used to magnetize the tissues.
- Radiofrequency is applied to tip magnetization out of equilibrium states.
- Gradients are added to encode the spatial origin of the signal.
- Finally, the acquired signal requires a Fourier transform to obtain the final image.
- Images can be sensitive to T1 and T2 relaxation of the tissues by appropriately tuning TE and TR of the sequence.
- Two main type of sequences are the spin echo and the gradient echo sequences.
- Contrast agents can be used to enhance pathology visualization; they are mainly Gadolinium-based.

## Chapter Outline

**MRI System**

**MRI Principle**

**The Spin Echo Sequence**

**The Gradient  
Echo Sequence**

**Other Sequences: The  
MRI Jungle**

**Contrast Agents in MRI**

**Artefacts in MRI**

**MRI Advantages  
and Disadvantages**

▶ **Take Home Messages**

**References**

**Test Your Knowledge**



## References



Excellent websites to understand the MRI technique and all its related questions :

- <https://www.imaios.com/en/e-mri>
- <https://mrquestions.com/index.html>

### MRI safety :

A Practical Guide to MR Imaging Safety: What Radiologists Need to Know

Leo L. Tsai, Aaron K. Grant, Koenraad J. Mortelet, Justin W. Kung, and Martin P. Smith

RadioGraphics 2015 35:6, 1722-1737

### MRI physics :

Plewes, D.B. and Kucharczyk, W. (2012), Physics of MRI: A primer. J. Magn. Reson. Imaging, 35: 1038-1054. <https://doi.org/10.1002/jmri.23642>

### Sequences:

Jung, B.A. and Weigel, M. (2013), Spin echo magnetic resonance imaging. J. Magn. Reson. Imaging, 37: 805-817. <https://doi.org/10.1002/jmri.24068>

MR Pulse Sequences: What Every Radiologist Wants to Know but Is Afraid to Ask

Richard Bitar, General Leung, Richard Perng, Sameh Tadros, Alan R. Moody, Josee Sarrazin, Caitlin McGregor, Monique Christakis, Sean Symons, Andrew Nelson, and Timothy P. Roberts

RadioGraphics 2006 26:2, 513-537

## Chapter Outline

**MRI System**

**MRI Principle**

**The Spin Echo Sequence**

**The Gradient  
Echo Sequence**

**Other Sequences: The  
MRI Jungle**

**Contrast Agents in MRI**

**Artefacts in MRI**

**MRI Advantages  
and Disadvantages**

**Take Home Messages**

► **References**

**Test Your Knowledge**



## Test Your Knowledge



Which of the following statements are correct ?

1- Some objects will be attracted by an uncontrollable force into the MRI scanner bore due to the main magnetic field; these are :

- a) Ferromagnetic objects (Iron, nickel, cobalt and their alloys)
- b) Metallic objects (all that are electrically conductive)
- c) All medical implants without exception

2- Which element of the MRI system allows encoding spatial origin of the signal emitted :

- a) Main magnet
- b) Radiofrequency
- c) Gradients x,y,z

3- The magnetization of the tissues occurs when :

- a) The MRI sequence is starting
- b) The subject receives radiofrequency wave
- c) The subject is lying on the table inside the scanner bore

4- The echo time TE is the time between :

- a) The RF excitation pulse and the RF refocusing pulse in the spin echo sequence
- b) The RF excitation pulse and the echo emission in the gradient echo sequence
- c) Two consecutive RF excitation pulses

## Chapter Outline

**MRI System**

**MRI Principle**

**The Spin Echo Sequence**

**The Gradient  
Echo Sequence**

**Other Sequences: The  
MRI Jungle**

**Contrast Agents in MRI**

**Artefacts in MRI**

**MRI Advantages  
and Disadvantages**

**Take Home Messages**

**References**

► **Test Your Knowledge**



### Test Your Knowledge



Which of the following statements are correct ?

=> **ANSWERS**

1- Some objects will be attracted by an uncontrollable force into the MRI scanner bore due to the main magnetic field; these are :

- a) **Ferromagnetic objects (Iron, nickel, cobalt and their alloys)**
- b) Metallic objects (all that are electrically conductive)
- c) All medical implants without exception

2- Which element of the MRI system allows encoding spatial origin of the signal emitted :

- a) Main magnet
- b) Radiofrequency
- c) **Gradients x,y,z**

3- The magnetization of the tissues occurs when :

- a) The MRI sequence is starting
- b) The subject receives radiofrequency wave
- c) **The subject is lying on the table inside the scanner bore**

4- The echo time TE is the time between :

- a) The RF excitation pulse and the RF refocusing pulse in the spin echo sequence
- b) **The RF excitation pulse and the echo emission in the gradient echo sequence**
- c) Two consecutive RF excitation pulses

### Chapter Outline

**MRI System**

**MRI Principle**

**The Spin Echo Sequence**

**The Gradient  
Echo Sequence**

**Other Sequences: The  
MRI Jungle**

**Contrast Agents in MRI**

**Artefacts in MRI**

**MRI Advantages  
and Disadvantages**

**Take Home Messages**

**References**

► **Test Your Knowledge**





## Test Your Knowledge



5- The repetition time TR is the time between :

- a) The RF excitation pulse and the RF refocusing pulse in the spin echo sequence
- b) The RF excitation pulse and the echo emission in the gradient echo sequence
- c) Two consecutive RF excitation pulses

6- T1-weighted contrast is obtained with

- a) A short TE and a short TR
- b) A long TE and a short TR
- c) A long TE and a long TR

7- For a T2-weighted image, a long TR is used so that the magnetization is regrowing to its initial state at equilibrium between each successive RF pulse. How should the TE be?

- a) Short TE
- b) Long TE
- c) TE should equal TR/2

8- Which sentence is correct :

- a) An MRI exam is cheap and fast
- b) An MRI exam is long and more expensive than CT
- c) An MRI exam is very quiet

9- With an MRI I can get :

- a) Excellent soft tissue contrast but no other information
- b) Anatomical images, information about water diffusion, parameters related to brain activation for motor task
- c) Excellent bone contrast and poor discrimination of soft tissues

## Chapter Outline

**MRI System**

**MRI Principle**

**The Spin Echo Sequence**

**The Gradient  
Echo Sequence**

**Other Sequences: The  
MRI Jungle**

**Contrast Agents in MRI**

**Artefacts in MRI**

**MRI Advantages  
and Disadvantages**

**Take Home Messages**

**References**

► **Test Your Knowledge**



### Test Your Knowledge



=> ANSWERS

5- The repetition time TR is the time between :

- a) The RF excitation pulse and the RF refocusing pulse in the spin echo sequence
- b) The RF excitation pulse and the echo emission in the gradient echo sequence
- c) **Two consecutive RF excitation pulses**

6- T1-weighted contrast is obtained with

- a) **A short TE and a short TR**
- b) A long TE and a short TR
- c) A long TE and a long TR

7- For a T2-weighted image, a long TR is used so that the magnetization is regrowing to its initial state at equilibrium between each successive RF pulse. How should the TE be?

- a) Short TE
- b) **Long TE**
- c) TE should equal TR/2

8- Which sentence is correct :

- a) An MRI exam is cheap and fast
- b) **An MRI exam is long and more expensive than CT**
- c) An MRI exam is very quiet

9- With an MRI I can get :

- a) Excellent soft tissue contrast but no other information
- b) **Anatomical images, information about water diffusion, parameters related to brain activation for motor task**
- c) Excellent bone contrast and poor discrimination of soft tissues

### Chapter Outline

**MRI System**

**MRI Principle**

**The Spin Echo Sequence**

**The Gradient Echo Sequence**

**Other Sequences: The MRI Jungle**

**Contrast Agents in MRI**

**Artefacts in MRI**

**MRI Advantages and Disadvantages**

**Take Home Messages**

**References**

► **Test Your Knowledge**



All material used (including intellectual property and illustration elements) either originates from the authors, the authors were entitled to use the material by applicable law or have obtained a transferable license from the copyright holder.

## Chapter Outline

**MRI System**

**MRI Principle**

**The Spin Echo Sequence**

**The Gradient  
Echo Sequence**

**Other Sequences: The  
MRI Jungle**

**Contrast Agents in MRI**

**Artefacts in MRI**

**MRI Advantages  
and Disadvantages**

**Take Home Messages**

**References**

**Test Your Knowledge**